

GL SpectroSoft Report  
Lighting Research & Measurement Center

Measurement name: B3884\_F\_20231129\_R\_S  
Measurement time: 2023-11-29T16:37:44  
Report time: 2023-11-30  
Device: GL SPECTIS 1.0 Touch + Flicker  
Sensor Temperature: 28.16°C

Client name: LED line LITE  
Model name: Plafon 24W 4000K 1950lm IP44  
z czujnikiem ruchu  
Item No: 201446  
Serial No.: 22005  
Made by: Radosław Kośnik

## Flicker

Flicker	100 Hz
Flicker index	0.00141
Flicker percent	1.87 %
SVM	0.009

## Flicker chart



## FFT chart

